

台中捐血中心
血小板細菌檢測累計彙整報告

月份	抽檢人次	初測陽性數 (initial positive)	偽陽性 數	偽陽性率 %	確認陽 性數	確認陽性 率%	偽陰性 數	偽陰性率 %	確認陽性/偽陰性鑑定結果
97.01~12	9,746		26	0.27	2	0.02	0	0	1.None found 2.Citrobacter koseri
98.01~12	11,186		11	0.10	0	0	0	0	
99.01~12	14,591		4	0.03	0	0	0	0	
100.01~12	18,741		11	0.06	4	0.02	0	0	1.Streptococcus Bovis 2.Escherichia coli 3.Group D Enterococcus 4.Klebsiella pneumoniae
101.01~12	21,986		4	0.02	0	0	0	0	
102.01~12	24,062		16	0.07	3	0.01	0	0	1.Escherichia coli 2.Listeria monocytogenes 3.Coagulase (-) Staphylococcus
103.01~12	22,456		8	0.04	2	0.01	0	0	1.Streptococcus gallolyticus 2. Listeria monocytogenes
104.01~12	23,689		10	0.04	2	0.008	0	0	1.Streptococcus species 2.Streptococcus species
105.01~12	24,365		6	0.02	2	0.01	0	0	1. Streptococcus gallolyticus 2. Streptococcus gallolyticus
106.01	2,045	0	0	0	0	0	0	0	
106.02	2,066	0	0	0	0	0	0	0	
106.03	2,242	1	0	0	0	0	0	0	
106.04	2,310	1	0	0	0	0	0	0	
106.05	2,271	0	0	0	2	0.018	0	0	1.Klebsiella pneumoniae ssp 2.Staphylococcus aureus(MRSA)
106.06	2,342	1	0	0	0	0	0	0	
106.07	2,408	0	0	0	0	0	0	0	
106.08	2,387	0	0	0	0	0	0	0	
106.09	2,238	1	0	0	0	0	0	0	

備註:1. 初測陽性(initial positive)定義:培養曲線平直、平底未變色、菌種鑑定結果為 no growth，則判定為儀器因素造成之陽性。

2. 97~105 年初測陽性數(initial positive)併入偽陽性數計算。