

醫療財團法人台灣血液基金會台中捐血中心

血液成品 QC 測試報告

製備場所：成分課

測試月份：108 年 05 月

Components		Test	No of tested /No of prepared	Test Result		Target Value
				Mean	SD	
紅血球類	Packed RBCs	Hct	28/15,692	64	4.3	55-80% w/o SAGM
			NA	NA	NA	50-70% w. SAGM
	Leukocyte Reduced RBCs	leukocyte count (×10 ⁶ /unit)	29/25,053	0.019	0.0093	<5×10 ⁶ /unit
血小板類	Platelet Concentrate	platelet count (×10 ¹⁰ /unit)	24/570	6	1.3	≥2.75×10 ¹⁰ /unit
		pH	24/570	7	0.27	≥6.2
	Apheresis Platelets	platelet count (×10 ¹¹ /unit)	6/322	3.6	0.28	≥3.0×10 ¹¹ /unit
		pH	6/322	7.5	0.02	≥6.2
	Leukocyte Reduced Apheresis Platelets	leukocyte count (×10 ⁶ /unit)	30/4,104	0.095	0.088	<5×10 ⁶ /unit
		platelet count (×10 ¹¹ /unit)	30/4,104	3.6	0.56	≥3.0×10 ¹¹ /unit
		pH	30/4,104	7.2	0.25	≥6.2
	血漿類	FFP	Factor VIII	13/25,021	72	26
RBC count (×10 ⁹ /L)			30/25,021	0.21	0.27	<6.0×10 ⁹ /L
leukocyte count (×10 ⁹ /L)			30/25,021	0.01	0.01	<0.1×10 ⁹ /L
platelet count (×10 ⁹ /L)			30/25,021	11	9.8	<50×10 ⁹ /L
Cryo		factor VIII	12/10,488	67		≥40 IU /unit
		fibrinogen	12/10,488	262	54	≥75 mg /unit

備註：Factor VIII、fibrinogen 測試頻率為每 2 個月執行。

表 QP-BB-032-6.7-001